



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Condition Report

Sale 1826 - 11/12/2018 10:00:00 - Medical Equipment Auction - PART 2

Sale Date 11-12-2018 10:00

Location: Malmesbury. Wiltshire, UK

Viewing: Friday 7th December 12pm - 5pm, Monday 10th December 10am - 6pm and the morning of the sale from 8:30am

Removal: By 4pm on Tuesday 18th December 2018 - Late collection charges will be incurred after this date

These reports are compiled by our Medical Engineering Team and are for guidance only - all lots are sold as seen, as per our Conditions of Sale. We recommend that wherever possible you inspect lots prior to purchase.

Lot 803 Datex Ohmeda Aisys Anaesthetic Machine with 3 x Vaporisers, 2 x Flat Screen

Software: 08.01

Condition Report: The anaesthetic machine powers up and completes the initial system self test successfully. Display screens and control panels are in good working order. Both the ventilator and the CareScape patient monitoring system appear to be functional but have not been fully tested. The breathing system is missing the flow sensors but the flow sensor module, O2 sensor and bellows are in place (CO2 absorber canister is not included). The backup battery found to be working at the time of test. Air, O2, N2O, Suction and AGSS hoses are included. Overall the unit is in reasonable physical condition.

Lot 804 PFI Datex Ohmeda Aisys Anaesthetic Machine with 3 x Vaporisers, 2 x Flat

Software: 08.01

Condition Report: The anaesthetic machine powers up and completes the initial system self test successfully. Display screens and control panels are in good working order. The ventilator appears to be functional but has not been fully tested. The CareScape patient monitoring system powers up and starts the initialisation process but fails to proceed to the main screen (automatically re-starts continuously). The breathing system is missing the flow sensors but the flow sensor module, O2 sensor and bellows are in place (CO2 absorber canister is not included). The backup battery found to be working at the time of test. Air, O2, N2O, Suction and AGSS hoses are included. Overall the unit is in reasonable physical condition.



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Lot 835 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE84384303 & DE84382998

Manufactured: 02 2009

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 855 Topcon Ophthalmometer and Handheld Keratometer

Condition Report: The Topcon OM-4 ophthalmometer powers up, lamp and brightness control are in good working order and the optics appears to be clear (not fully tested). The joystick control, chin rest adjustment and table height adjustment are all functional. The Nidek KM-500 auto keratometer also powered up and the display panel/function keys were operational but the battery was low and no further checks could be carried out. Charger is included. Overall the unit is in reasonable physical condition.

Lot 857 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE84382999 & DE84383001

Manufactured: 2009

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 858 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE84384318 & DE84382997

Manufactured: 2009

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 859 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE84384325 & DE4323

Manufactured: 2009

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

was not fully tested. Lot are in good physical condition.

Lot 863 Carefusion Microlab Spirometer

Serial No: 085-67729

Manufactured: 2012

Associated Equipment: External printer connection cable
USB connection cable

Accessories: User manual
Spirometer attachment
Disposable spirometer tubes
Noseclip
Hard plastic carry case with foam inserts
Power supply & cable

Software: V2.43

Condition Report: Equipment powers-up and completes self-test successfully. Display, touch-screen and softkeys in satisfactory working condition. Spirometer attachment connectivity and function responsive and satisfactory. Printer output fully-functional. Good physical condition.

Lot 873 Haag Streit Slit Lamp

Condition Report: The slit lamp powers up and the lamp and brightness control are in good working order. The binocular assembly is intact with eyepieces in place and the system produces clear images. The joystick control is functional but the roller wheel on one side of the slit lamp is broken. The chin rest adjustment and table height adjustment are in good working order. Overall the unit is in satisfactory physical condition. The mains cable for the power supply is slightly loose causing intermittent power loss.

Lot 947 Eschmann TD830 Diathermy/Footswitch | DOM 10/2006 | with Smoke Evacuator

Condition Report: The electrosurgical unit powers up and completes the initial system self test successfully. Display panel, controls and rotary power selectors are in good working order. The patient plate alarm and indicator are functional, both bipolar and monopolar modes are working. Tested with electrosurgical analyser and the output power from all modes (cut, coagulation and bipolar) were within the limits at the time of test. Footswitches are included and functional. The smoke evacuator powers up ok but was not fully tested. Overall the units are in satisfactory physical condition with minor signs of wear.

Lot 952 2 x Philips MP70 Intellivue Patient Monitors

Serial No: DE84383003 & DE84382996

Manufactured: 2009



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 1014 Draeger 5400 Transport Incubator with Draeger Babylog 2000 Ventilator | 117325 |

Serial No: ARTB-0009

Manufactured: 2003

Accessories: Air & O2 hose
Patient circuit x3
Draeger Oxydig O2 monitor
Draeger Babylog 2000 with Exp. valve

Condition Report: The incubator power up on mains and the set air temperature is maintained accordingly. The Babylog ventilator was tested with medical gases and test lung. Good ventilation cycles were delivered. The parameter controls, displays, indicators and alarm settings are in good working order. The air/O2 blender and O2 therapy flow meter are fully functional. The gas operated low suction system is working fine. The Oxydig O2 monitor needs new batteries. System is in good overall condition.

Lot 1017 Haag Streit Slit Lamp

Condition Report: The slit lamp powers up but the lamp is missing. The binocular assembly is intact with eyepieces in place and the optics appear to be clear (not fully tested). The fixation light, joystick control, chin rest adjustment and table height adjustment are in good working order. The base plate of the storage box is missing and the lid is also loose. Otherwise the unit is in satisfactory physical condition with signs of wear.

Lot 1036 Agfa Model CR35X CR Reader

Serial No: 7923

Manufactured: 10 2011

Accessories: None

Software: c25_3206

Condition Report: Equipment powers-up and completes self-test successfully. Display and softkeys in satisfactory working condition. Not fully tested. Good physical condition.

Lot 1038 Whitestar Signature Phaco Machine with Footswitch

Serial No: 200851220

Manufactured: 10 2008

Accessories: Footswitch

Software: V6.023

Condition Report: Equipment powers-up and completes self-test successfully. Display and touch-screen in satisfactory working condition. Footswitch currently untested, but appears to be in good



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

physical condition. Not fully tested. Good overall physical condition.

Lot 1067 Philips Sonos 7500 Ultrasound Machine with Philips S4 Probe

Serial No: US60421019

Manufactured: 07 2004

Associated Equipment: Philips S4 Cardiac Probe model No. 21330A, S/N: US97313634 912/2002)

Software: D.2

Condition Report: The Sonos 7500 power up and completed the system initialisation. Control console's trackball, gain knobs, user interface touch screens, CRT monitor and keys are in good working order. The cardiac probe was tested with a test phantom, good image quality was displayed with fair resolution and grey scale. Calipper measures were accurate. Available modes: 2d, M-mode, Colour, PW, CW & Angio. Good ECG trace displayed when tested with ECG lead and simulator. Patient registration & storage are functional but there MO disk is not included. Unit is in a fair overall condition.

Lot 1077 Eschmann Little Sister SES 113B Autoclave

Manufactured: 2010

Condition Report: The autoclave powers up and completes the initial system self test successfully. Display screen, control panel and printer are in good working order and the door locking mechanism is also functional. The instrument rack is in place but the unit does not come with any trays. Water containers are included. No further checks carried out. During the start up the system displays 'service required' but the unit is usable. Mains power cable is not included. Overall the unit is in reasonable physical condition. Cycle count: 12063.

Lot 1078 Eschmann Little Sister SES 113B Autoclave

Manufactured: 2011

Condition Report: The autoclave powers up and completes the initial system self test successfully. Display screen, control panel and printer are in good working order and the door locking mechanism is also functional. The instrument rack is in place but the unit does not come with any trays. Water containers are included. No further checks carried out. During the start up the system displays 'service required' but the unit is usable. Mains power cable is not included. Overall the unit is in reasonable physical condition. Cycle count: 11448.

Lot 1110 Datex Ohmeda S/5 Avance Anaesthetic Machine

Condition Report: The anaesthetic machine powers up and completes the initial system self test



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

successfully. Display screen and control panel are in good working order. The ventilator appears to be functional but has not been fully tested. The breathing system is intact with flow sensors, O2 sensor and bellows in place (single use CO2 absorber canister is also included). The backup battery found to be working at the time of test. Medical gas hoses are not included. Overall the unit is in satisfactory physical condition.

Lot 1115 Datascope CS100 Intra-Aortic Balloon Pump

Operating Hours: 904

Condition Report: The unit powers up and completes the initial system self test successfully. LCD display screen, control panel, pump and printer are in good working order. The adult condensate removal / safety disk is in place (expiration date Jan 2019) but the Helium gas cylinder is not included. The unit does not come with any accessories. No further checks carried out. Overall the unit is in reasonable physical condition. The power cable holding bracket is damaged.

Lot 1123 Wild Heerbrugg Leica Microscope

Condition Report: The M650 operating microscope powers up and the lamp and brightness control are in good working order. The optics appear to be clear but the unit does not come with a binocular to carry out full test. Focus and magnification controls are functional, objective lens (250mm) is included and overall the unit is in reasonable physical condition.

Lot 1241 GE Mac5000 ECG

Serial No: AAY05055384PA

Manufactured: Est. 2004

Accessories: ECG leads and connection

Condition Report: Powers-up and performs self-test successfully. Display and softkeys in satisfactory working condition. Not fully tested. Good physical condition.

Lot 1242 GE Mac5000 ECG

Serial No: AAY04514753PA

Manufactured: Est. 2004

Accessories: None

Condition Report: Powers-up and performs self-test successfully. Display and softkeys in satisfactory working condition. Not fully tested. Good physical condition.

Lot 1253 Haag Streit 900 Slit Lamp



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Condition Report: The slit lamp powers up and the lamp and brightness control are in good working order. The binocular assembly is intact but the eyepieces are missing. The joystick control is functional but the roller wheel on one side of the slit lamp is broken and the covers for the roller wheels/rails are missing. The chin rest adjustment and table height adjustment are in good working order. Overall the unit is in reasonable physical condition.

Lot 1258 Puritan Bennett 840 Ventilator

Operating Hours: 16647

Condition Report: The ventilator powers up and completes the initial self test successfully. Display panel, touch screen and control panel are in good working order (touch screen is intermittently unresponsive). All ventilation modes are selectable but the system has not been fully tested. Neonatal mode is not available on this unit. The backup battery found to be working at the time of test. Air and O2 hoses are included. Overall the unit is in reasonable physical condition. Operating hours: 16647.

Lot 1259 Puritan Bennett 840 Ventilator

Operating Hours: 24557

Condition Report: The ventilator powers up and completes the initial self test successfully. Display panel, touch screen and control panel are in good working order. All ventilation modes are selectable but the system has not been fully tested. Neonatal mode is not available on this unit. The backup battery found to be working at the time of test. Air and O2 hoses are included. Overall the unit is in reasonable physical condition. Operating hours: 24557.

Lot 1264 Philips ECG Stress Tester and Suntech Tango

Serial No: CZC91356JL

Manufactured: Est. 2007

Accessories: HP stack unit CZC91356JL
StressVue thermal printer 106031745163
StressVue network dongle
SunTech Tango+ monitor 26149

Condition Report: Equipment powers-up successfully. Display and softkeys in satisfactory working condition. Problem with CMOS battery in stack unit - triggers 'boot setup' option on self-test. Also paper-jam present in printer module. Not fully tested. Good overall physical condition.

Lot 1310 Conmed HD DRSHD Tower

Serial No: DRSHD-1213-021C

Manufactured: Est. 2005



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Software: V5.5.3.8

Condition Report: Powers-up and completes self-test successfully. Software stores patient images and data as expected. Good physical condition.

Lot 1312 Medxchange HD Computer Tower

Serial No: DRSHD-0810-036C

Condition Report: No hard drive

Lot 1376 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE61754671 & DE35017097

Software: Rev G.01.80

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition. Date of manufacture 04/2007 & 03/2004.

Lot 1377 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE35017091 & DE61754713

Software: Rev G.01.80

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition. Date of manufacture 04/2007 & 03/2004.

Lot 1378 2 x Philips Intellivue MP70 Patient Monitors

Serial No: DE61755892 & DE61754897

Manufactured: 2007

Software: Rev G.01.72 & G.01.80

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 1427 Nicolet Biomedical Viking Select EMG Unit with Accessories and HP Deskjet

Condition Report: Hard drive formatted.

Lot 1512 LX Strobe Larynograph with Sony Monitor (No HDD) | SN 6005961

Condition Report: Hard Drive removed

Lot 1593 Topcon Slit Lamp, Table and Power Supply



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Condition Report: The Topcon SL-7F slit lamp powers up and the lamp and brightness control are in good working order. The binocular assembly is intact with eyepieces in place and the system produces clear images. The magnification selector knob, joystick control, chin rest adjustment and motorised table height adjustment are in good working order. The Topcon FD-12E flash device also powers up and appears to be functional but the lamp is not working. Overall the unit is in good physical condition.

Lot 1602 Haag Streit BQ900 Slit Lamp On Table (No Eye Piece Section)

Condition Report: The slit lamp powers up but the main bulb appears to be not working (fixation light is in good working order). The optics appear to be clear but the binocular is not included to carry out further checks. The magnification selector knob, joystick control, chin rest adjustment and table height adjustment are in good working order. Overall the unit is in good physical condition.

Lot 1627 Micropace Cardiac Stimulator | 0739, Bona Computech Light System and Monitor

Serial No: MP3093-W0-07/020

Manufactured: Est. 2007

Accessories: MP3093 Bona Computech light system MP3093-W0-07/020
MP3016 Computer keyboard MC0410-07/02-011
Elo LCD touch-screen monitor
Bona PC power supply and cable

Software: V3.2

Condition Report: Powers-up and completes self-test successfully. Display, and softkeys in satisfactory working condition. Stylus input not working. Touch-screen untested, due to lack of working stylus input, but appears to be in good physical condition. Not fully tested. Good overall physical condition.

Lot 1653 Leica DMLS Microscope

Condition Report: The lab microscope powers up and the lamp and brightness control are in good working order. The optics appear to be clear but the unit does not come with any objective lenses. Eyepieces are included. Stage movement with coarse and fine adjustment controls are in good working order but the stage clip/test slide holder is missing. Light condenser is in place and overall the unit is in satisfactory physical condition.

Lot 1654 Leica Microscope

Condition Report: The lab microscope powers up and the lamp and brightness control are in good working order. The optics appear to be clear but the unit does not come with any objective lenses. Eyepieces are included. The coarse and fine adjustment controls appear to be



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

working. Overall the unit is in satisfactory physical condition.

Lot 1694 Keeler Ophthalmoscope

Condition Report: The ophthalmoscope does not power up. Battery and bulb are both in place (battery might be depleted). The unit has sustained damage to the plastic frame close to the lens.

Lot 1707 Barco Monitor, HP PC Tower, Eaton UPS On Agfa Medical

Condition Report: The unit powers up but HDDs have been formatted.

Lot 1721 Olympic CFM 6000 Cerebral Function Monitor on Stand

Serial No: 11150

Manufactured: 2011

Accessories: CFM 6000 Cerebral function monitor on stand 11150
CFM Model 6000 Amplifier module A7299
EEG Connection leads

Condition Report: Powers-up and completes self-test successfully. Display and touch-screen in satisfactory working condition. Function menus fully accessible. Not fully tested. Good physical condition.

Lot 1722 Olympic CFM6000 Cerebral Function Monitor on Stand

Serial No: 10594

Manufactured: 2011

Accessories: CFM6000 Cerebral function monitor on stand 10594
CFM Model 6000 Amplifier module A-8071
EEG Connection leads

Condition Report: Powers-up and completes self-test successfully. Display and touch-screen in satisfactory working condition. Function menus fully accessible. Not fully tested. Good physical condition.
